




# Apex locator Woodpex III

## Instruction Manual



Please read this manual before operating.

Industrial design patent No.: CN 201130379058.9

**Guilin Woodpecker Medical Instrument Co., Ltd.**

# Contents

1 Introduction .....	1
2 Notice of installing and using the device .....	5
3 Installation of the device .....	7
4 Product function and operation .....	11
5 Trouble shooting .....	17
6 Disinfection, cleaning and sterilization procedure for the accessories ....	20
7 Storage, maintenance and transportation .....	23
8 Environmental protection .....	24
9 After service .....	24
10 Symbol instruction .....	25
11 European authorized representative .....	26
12 Statement .....	26
13 EMC - Declaration of conformity.....	26

# **1 Introduction**

## 1.1 Foreword

Guilin Woodpecker Medical Instrument Co., Ltd. is a professional manufacturer in researching, developing and producing dental equipment which has a wholesome quality assurance system.

There are two brands of WOODPECKER Company: WOODPECKER and DTE. Products include ultrasonic scaler, curing light, apex locator and ultrasurgery, etc.

## 1.2 Description of the device

Apex locator is a supporting equipment of endodontic treatment, through the measurement of the length of apical teeth, helping dentists to finish the endodontic treatment.

Features of the device:

- a) Equipped with clear bright LCD, clear image and different colors indicate the trajectory of the file clearly.
- b) Based on advanced multiple frequency network impedance measurement technology and automatic calibrating ensures the measurements are accurate.
- c) The File clip, Lip hook and Touch probe can be autoclaved under high temperature and high pressure. Avoiding cross infection effectively.
- d) Battery is rechargeable, unnecessary to replace batteries repeatedly.

### 1.3 Model and dimensions


1.3.1 Model: Woodpex III

1.3.2 Dimensions: 84mm (length) × 88mm (width) × 112mm (height)

1.3.3 Weight: 336g


### 1.4 Components

1.4.1 Picture of the main unit. ( Picture 1 )



Picture 1

### 1.4.2 Pictures of the main accessories ( Picture 2 )



### 1.5 Structure

Woodpex III is composed of main unit, measuring wire, lip hooks, file clip, touch probe, adapter, etc..

### 1.6 Intended use

This equipment applies to the measurements below:

1.6.1 Measurement of pulpitis, pulp necrosis, periapical periodontitis and tooth length.

1.6.2 Measurement of the tooth length before restoration of post crown.

1.6.3 Measurement of the tooth length of transplantation and retransplantation.

## 1.7 Contraindication

We do not advise the use of Woodpex III on patients fitted with pacemakers (or other electrical equipment) or on those patients who are advised not to use the electric equipment (like electric shaver, electric blower) for safety reasons.

## 1.8 The classification of the device

1.8.1 Type of protection against electric shock: Class II equipment

1.8.2 Degree of protection against electric shock: Type BF applied part

1.8.3 Degree of protection against water shock: Ordinary equipment ( IPX0 )

1.8.4 Device not suitable for being used in the presence of a flammable anesthetic mixture with air or with oxygen or nitrous oxide.

1.8.5 Operation mode: Continuous operation

## 1.9 The main technical specifications

1.9.1 Battery: 3.7V/750mAh

1.9.2 Adapter: ~100V-240V 50Hz/60Hz

1.9.3 Consumption power:  $\leq 0.5W$

1.9.4 Screen: 4.5" LCD

1.9.5 Buzzer alert: The buzzer will alert when the endo file is less than 2mm to

the apex.

#### 1.9.6 Operation condition

- a) Environment temperature: +5°C ~ +40°C
- b) Relative humidity: 30% ~ 75%.
- c) Atmosphere pressure: 70kPa ~ 106kPa

## 2 Notice of installing and using the device



2.1 Please read the instruction manual carefully before the operation.

2.2 As a safety precaution in order to avoid over-instrumentation, it is recommended to proceed as follows: place the file onto an endodontic ruler, where the apex locator screen indicates '00'. Subtract 0.5-1 mm from the measured file length as the Working Length.

2.3 The scale indication on the apex locator screen does not represent a distinct length or distance in mm or other linear units. It simply indicates the touch probe progression towards the apex foramen.

2.4 If the screen bar graph suddenly makes a large movement or immediate display 'OVER' in the upper part of the canal, continue slightly towards the apex so the signal returns to normal.

2.5 In order to prevent leakage or interference between the root canal and resulting in inaccurate measurements, dry the access cavity with a cotton pellet or air-blower before each use.

- 2.6 Use a file size adapted to the root canal diameter. The selected file is too small for a large root canal might cause the screen digital display is not steady during the procedure.
- 2.7 In order to confirm the file clip and measuring wire makes good contact, test the wire connecting before each use(See 3.1.3).
- 2.8 The file clip, lip hook and touch probe are reusable. Please make sure they are autoclaved under high pressure and high temperature before each operation. The endo files should not be used more than 3 times.
- 2.9 The batteries must be taken out for storage when the device is not used for a long time.
- 2.10 Please recharge the battery when low battery indicator flashes.
- 2.11 Please use original components, the components made by other companies may cause inaccurate measurement or un-measurable.
- 2.12 Avoid the connection between the outside and inside liquid of endodontic during measuring in order to avoid the measuring difference.
- 2.13 Keep endo file and file clip away from any other metal or instruments.
- 2.14 To ensure that short circuits do not impair the measurements, be particularly careful with patients fitted with metal crowns or bridges. Please confirm the wetness of the endo to ensure the reliability of the measuring. If it is confirmed that the endo file hasn't reached the apex yet the data showed on the apex locator is too low, please check whether the endo is too dry and confirm it with X-ray.
- 2.15 This device have electromagnetic interference, the patient or doctor who with

a heart pace maker are forbidden to use this device and the device is susceptible to other device which produces electromagnetic interference. Dentists should be cautious about operation under such environment.

2.16 The guarantee is valid for normal usage conditions. Any disassembly will render the guarantee void, the professionals of Woodpecker company will offer the repair service during guarantee period.

2.17 Any modification will render the guarantee void and may cause harm to the patient.

2.18 Only the original adapter and lithium battery could be used to this machine.

### **3 Installation of the device**

#### 3.1 The Connection of the Measuring Wire

3.1.1 Insert the plug of the measuring wire into the right side socket of the unit.

**Attention:**


a) Please be careful to use the device, keep it stable and avoid hit. Incautious use will lead to the damage or the failure of the machine.

b) Measurement can not be proceeded without the complete insertion of the plug.


c) Be sure not to hit the plug. Keep the device away.

3.1.2 Insert the file clip and lip hook respectively into the two sockets of the measuring wire. [Picture 3]


**Attention:**



Picture 3



Picture 4



Be sure not to pull the wire when inserting or pulling out the the measuring wire and the file clip. [Picture 4 (a)]


Correct operation showed as in picture 4 (b).

### 3.1.3 Test the wire connecting (Test before each use)


a) Press the power switch. Make sure the scene of measuring the length of the root canal displayed on the LCD screen.

The device will shutdown automatically after 5 minutes without operation. [Picture 5]

b) Make sure if the plug of the measuring wire is inserted into the socket



Picture 5




Picture 6


correctly.

c) Make sure if the file clip and lip hook are connected well to the measuring wire.


d) Make the lip hook touch the bent wire of the file clip [as showed in picture




(a)



(b)



(c)



(e)

- ① Approx 2mm to apical foramen
- ② Approx 0.6mm to apical foramen
- ③ Apex(apical foramen)

Picture 7

6] to confirm all the instruction bars are displayed on the LCD screen and static display the digital '-3', otherwise , it means that the file clip or the measuring wire is damaged, should be replaced.[as showed in picture 7 (c)]

#### 3.1.4 Explanations on the interfaces displayed

- a) The screen displays the front region of the apical foramen by instruction bars. Please refer to the white region as showed. [Picture 7(a)]
- b) The file has gone to the position near by the apical foramen when the green bars displayed [Picture 7 (b)].
- c) The file has exceeded the apical foramen when the red bars displayed. A continuous beep sound will be generated at the same time [Picture 7 (c)].

#### 3.1.5 Enlarged display of apical foramen. [Picture 7 (d)]

### 3.2 Testing the device by tester(Test every two weeks)

Users can use the tester to check if the device work properly, specific operation is as follows:

- a) Pulling out the the measuring wire and turn off the device.
- b) Insert the tester.
- c) Turn on the device, it works fine when green area of the display light. (Digital display as 02, 03 or 04)

### 3.3 Demonstrate function

Demonstrate function can show the movement process of endo file when measuring. Specific operation is as follows:

- a) Pulling out the the measuring wire.

b) Turn on the device.

c) Long pressing the play button for 1 second can enter the demonstrate function which shows the movement process of endo file.

d) Repress the demonstrate function button then exit demonstrate function.

### 3.4 Charge the battery

When the power indicator flashes, please stop using the device, as it shows that the remaining power is not enough.

3.4.1 Connect the AC adapter and the socket on the left of the device properly, and insert the AC adapter plug to the power socket.

3.4.2 When the yellow indicator on, it shows the device is charging; when the yellow indicator turns to green indicator, it shows charging has finished. It may take 120minutes for charging.

3.4.3 After charging, please pull the AC adapter and plug out.

**Attention:** please don't use the device when it is charging, and the operator should be over 2meters away from the device.

## 4 Product function and operation

### 4.1 Usage requirements

Apex locator should be precise, repeatable, and easy to operate. The following requirements are necessary besides the proper operation method.

4.1.1 The operation should be according to the manual.

4.1.2 The dentists should have the knowledge of teeth position and average length and the skill to operate the device.

4.1.3 An fully exposed access cavity to show the pulpal cabin.

4.1.4 A X-ray photo to show the whole length and root canal of the teeth.

4.1.5 The endo file should not be too big nor too small to avoid cutting through the apical foramen.

4.1.6 Mark an anatomized symbol on the diseased tooth and memorize it on the case history. This symbol should be marked on the health bridge or on the tooth filled integrated. The position of the mark should be on the incisal edge of the anterior tooth or on the spire of the molars. For those bridge that's broken obviously, this symbol should be on the tooth surface supported by the dentin instead of on the suspended enamel.

4.1.7 The acute inflammation surrounding the apex has been gone and the infected material has been cleaned. It is also necessary to get rid of the pulp and necrosis tissue.

4.1.8 The following cases are not suited for a normal measurement:

a) The size of the root similar to the size of apical foramen.

In this case, the measurement result of the length of the root canal will be shorter than its real because of the hypoplasia of the root [Picture 8].


b) Bleeding or the blood overflow from the apical foramen.

In this case, the blood will overflow from the root canal and reaches gingival that the blood and the gingival will be on a conducting state which will cause an


inaccurate result while measuring. The measurement can be continued when the bleeding is stopped [Picture 9].

c) The tooth crown is broken.


The tissue of the gingival may reach the cavity of the endo hole at the broken



Picture 8



Picture 9




Picture 10


point which will cause inaccuracy because of the electronic conduction. The measurement can be continued when the crown is fixed by gypsum or other insulators [Picture 10].

d) There is a crack on the tooth root.


In this case, the crack may cause the electric leakage which will affect the accuracy of measurement [Picture 11].



Picture 11



Picture 12



Picture 13

e) A retreatment to an endo which was filled with gutta-percha.


Clean the remaining material in the root canal and fill it with little normal saline before a measurement [Picture 12].

f) There is a metal crown which has connected to the gingival.

It will cause an inaccuracy when the endo file touches metal crown [Picture 13].

Sometimes, the results of the Apex Locator and X-rays do not meet each other, which is neither because the machine is not normal, nor the photo is incorrect taken. The actual position of the apical foramen is different from the anatomical one, it is very common that the apical foramen slightly to the side of the root canal crowns. In this case, according to the shooting angle as the following picture show, it will cause illusion that the front tip of the root canal haven't reached the canal tip. [Picture 14]

(Because of the angle of X-rays, sometimes it can't take photo of the apical foramen properly, so it can't show the accurate position of the apical foramen.)



Picture 14

## 4.2 Instruction

4.2.1 Insert the plug of measuring wire into the socket in the side of main unit. Turn it on. The battery is on the left of screen.

4.2.2 The equipment is in the normal condition. The equipment shuts down after 5 minutes without use.

4.2.3 The volume is adjustable. Please press the volume bottom for a setting.

4.2.4 Hang the lip hook on the lip, make sure it contact the oral mucosa as a reference electrode [Picture 15].


4.2.5 Clip the file with file clip, approach to the apex, then there will be continuous alarm when the distance is less than 2mm [Picture 16].

**Attention:**


a) When grip the root canal with a needle file, please grip the upper of the metal part (near the root canal at the needle handle). If you grip the lower part (blade or moving part), it will wear the metal part of the file folder and the resin part. [Picture 17]

b) When measuring the length of root canal, please don't use the metal needle file.


If you operate the device without the dentistry glove, it will cause leakage and the result



Picture 15



Picture 16




Picture 17

of measurement will be inaccurate. Therefore, please use the resin needle file and remember don't touch the metal part with finger.


c) Please don't use the worn file clip, and it will make the result of measurement inaccurate.

d) Please reference the [Picture 18 (a)] to grip the needle file. If as [Picture 18 (b)], it can't properly measure the length of the root canal due to the improper force, and the front of the root canal pin is easy to wear.

4.2.6 When the file reaches the apex, adjust the rubber piece set on the endo file to the reference point (incisal edge or fossa edge), then pull out the endo file, measure the length between the top of the file and the rubber piece, and this is the working length of the tooth. It also can be use with the touch probe instead of file clip, when it is inconvenient to measure the back teeth [Picture 19].



Picture 18



Picture 19

4.2.7 The components that touch body must be autoclaved under high temperature and high pressure. The shell and measuring wire should be cleaned by 75% alcohol.

**Attention:** Avoid the silk-screen when cleaning.

## 5 Trouble shooting

Problems	Possible cause	Solutions
No power and no signal on the screen after the power on.	1. If the battery is placed correctly? 2. If the battery with no power?	1. Re-install the battery. 2. Recharge the battery.
The length of the root canal cannot be measured.	1. If the measuring wire is connected correctly? 2. If the measuring wire is broken?	Confirm the measuring wire is plugged firmly, link the lip hook with the file clip to check if the measuring wire is broken.
No sound of alarm.	If the volume is set at “mute”?	Adjust the sound level.

Problems	Possible cause	Solutions
<p>Display not steady while measuring: the measurement result is rather longer or shorter; numerical display irregular.</p>	<p>If the connection between the lip hook and the oral mucosa is ok?</p>	<p>Make sure the lip hook has contacted the oral mucosa at a good position.</p>
	<p>Is there a blood/saliva overflowing, glued to the crown?</p>	<p>Blood, liquid overflow from the root canal, glued to the crown or the tooth neck, will cause short-circuit then cause the in-normal phenomena. Clean the blood and the liquid.</p>
	<p>If the root canal is filled with blood, liquid?</p>	<p>Once the endo needle contact the surface of the root canal which is filled with blood, liquid, it will display “OVER” immediately. In this case, push the needle to the apical root canal, then the display will be normal, you can measure the length of the root canal correctly.</p>
	<p>If there is liquid, scrap on the tooth surface?</p>	<p>Clean the tooth surface.</p>
	<p>If the endo needle contact the gums?</p>	<p>The LCD will display “OVER” if the endo needle contact the gums.</p>

Problems	Possible cause	Solutions
Display not steady while measuring; the measurement result is rather longer or shorter; numerical display irregular.	If there is still pulp in the root canal?	If there is much pulp left in the root canal, the root canal length can't be measured correctly.
	If the needle touched the metal repaired material?	Once the needle touched the metal repaired material, current measurement from the gums to the periodontal tissue loss, the screen will display "OVER".
	If the adjacent surface has caries?	Current measurement flow from caries of the adjacent surface to gums, then the root canal length can't be measured correctly.
	Whether there is collateral or the tooth root is broken?	Once the needle reached the collateral or the broken part of the tooth root, current measurement will overflow from periodontal ligament, it displays "OVER".
	Is it because in addition to the top pulp chamber, low tooth crown? Or there are residues left?	Use rubber dam to prevent the current flow to gums.

Problems	Possible cause	Solutions
Display not steady while measuring: the measurement result is rather longer or shorter; numerical display irregular.	Are there cysts apical?	If there has cysts, the length of root canal can't be measured accurately.
	Whether the file clip is not clean or broken?	Clean the file clip by alcohol, or replace it.
	Whether the measuring wire is broken or poor contact?	Contact the both end of the measuring wire directly, it displays “-3”.
The length measurement indicator only full display near narrow part of the apical.	Whether the root canal is occlusive?	The display will be normal after penetrating the narrow part of apical.
	If the root canal is too dry?	Wet the endo with Hydrogen peroxide or NaCl.
	If the endo file is too small for a large root canal?	Replace the current endo file with a larger one.

\* If all above measures do not work, please contact us.

## 6 Disinfection, cleaning and sterilization procedure for the accessories

### 6.1 Foreword

The lip hook, the file clip and the touch probe must be cleaned, disinfected and sterilized before each use to prevent any contamination. This concerns the first use as well as the subsequent uses.

## 6.2 General recommendations

- The user is responsible for the sterility of the product for the first cycle and each further usage as well as for the usage of damaged or dirty instruments.
- For your own safety, please wear personal protective equipment (gloves, glasses, mask).
- Use only disinfecting solution which is approved for their efficacy (VAH/DGHM-listing, CE marking, FDA approval)

## 6.3 Procedure for lip hook, the file clip and the touch probe

Operation	Operating mode	Warning
1. Pre-Disinfection or Decontamination	- Soak immediately just after usage all instruments in a disinfectant solution combined with proteolytic enzyme if possible.	- Follow instructions and observe concentrations and immersion time given by the manufacturer (an excessive concentration may cause corrosion or others defects on instruments). - The disinfectant solution should be aldehyde free (to avoid blood impurities fixation). - Do not use disinfectant solution containing Phenol or any products which are not compatible with the instruments (See general recommendation). - For visible impurities that are observed on instruments, a pre-cleaning is recommended by brushing them manually with soft material.
2. Rinsing	- Rinse manually and abundantly the accessories with current water	

Operation	Operating mode	Warning
3. Manual Cleaning	- clean manually the accessories with an adequate brush, preliminary soaked in a clean pre-disinfectant solution	- the file clip mechanism has to be activated during the cleaning process (press several times the push button) - no visible impurities should be observed on the accessories
4. Rinsing	See point # 2	
5. Disinfection	- Immerse the accessories in a disinfectant solution (bactericidal, virucidal, fungicidal, tuberculocidal and aldehyde free) according to the manufacturer recommendations	- Follow instructions and observe concentrations and time given by the manufacturer
6. Final rinsing	- See point # 2 - After rinsing, the accessories have to be dried.	
7. Inspection	- Inspect devices and sort out those with defects.	- Dirty instruments must be cleaned and disinfected again.
8. Packaging	- Pack the devices in "Sterilization pouches".	- Check the validity period of the pouch given by the manufacturer to determine the shelf life. - Use packaging which are resistant up to a temperature of 141°C and in accordance with EN ISO 11607.

Operation	Operating mode	Warning
9. Sterilization	- Steam sterilization at: 134°C during 18 min.	- The accessories (lip hook, file clip and touch probe) must be sterilized according to the packaging labeling. - Use fractionated vacuum or gravity (less preferred) autoclaves (according to EN 13060, EN 285). - Use validated sterilization procedure according to ISO 17665-1 - Respect maintenance procedure of the autoclave device given by the manufacturer. - Use only the listed sterilization procedures.
10. Storage	- Keep devices in sterilization packaging in a dry and clean environment - Sterility cannot be guaranteed if packaging is open, damaged or wet (check the packaging before using the instruments).	

## 7 Storage, maintenance and transportation

### 7.1 Storage

7.1.1 This equipment should be stored in a room where the relative humidity is 10% ~ 93%, atmospheric pressure is 70kPa to 106kPa, and the temperature is -20°C ~ +55°C.

7.1.2 Avoid the storage in a too hot condition. High temperature will shorten the life of electronic components, damage battery, reshape or melt some plastic.

7.1.3 Avoid the storage in a too cold condition. Otherwise, when the temperature of the equipment increases to a normal level, there will be dew that will possibly

damage PCB board.

## 7.2 Maintenance

7.2.1 This device do not include accessories for repair usage, the repair should be carried out by authorized person or authorized after service center.

7.2.2 Keep the equipment in a dry storage condition.

7.2.3 Do not throw, beat or shock the equipment.

7.2.4 Do not smear the equipment with pigments.

## 7.3 Transportation

7.3.1 Excessive impact and shake should be prevented in transportation. Lay it carefully and lightly and don't invert it.

7.3.2 Don't put it together with dangerous goods during transportation.

7.3.3 Avoid solarization and getting wet in rain and snow during transportation.

## **8 Environmental protection**

Please dispose according to the local laws.

## **9 After service**

From the date this equipment has been sold, based on the warranty card, we will repair this equipment free of charge if there are quality problems. Please refer to the warranty card for the warranty period.

## 10 Symbol instruction



Trademark



Date of manufacture



Type BF applied part

IPX0

Ordinary equipment



Used indoor only



Power on / off



Sound adjustment



Humidity limitation



Demonstrate the measurement process



Appliance compliance WEEE directive



0197 CE marked product



Manufacturer



Class II equipment



Recovery



Keep dry



Handle with care



serial number



Temperature limitation



Atmospheric pressure for storage



Consult the accompanying documents



Authorised Representative in the EUROPEAN COMMUNITY

## 11. European authorized representative



MedNet GmbH  
Borkstrasse 10 · 48163 Muenster · Germany

## 12 Statement

All rights of modifying the product are reserved to the manufacturer without further notice. The pictures are only for reference. The final interpretation rights belong to GUILIN WOODPECKER MEDICAL INSTRUMENT CO., LTD. The industrial design, inner structure, etc, have claimed for several patents by WOODPECKER, any copy or fake product must undertake legal responsibilities.


## 13 EMC - Declaration of conformity

The device has been tested and homologated in accordance with EN 60601-1-2 for EMC. This does not guarantee in any way that this device will not be effected

by electromagnetic interference Avoid using the device in high electromagnetic environment.

<b>Guidance and manufacturer's declaration - electromagnetic emissions</b>		
The model Woodpex III is intended for use in the electromagnetic environment specified below. The customer or the user of the model Woodpex III should assure that it is used in such an environment.		
<b>Emissions test</b>	<b>Compliance</b>	<b>Electromagnetic environment - guidance</b>
RF emissions CISPR 11	Group 1	The model Woodpex III use RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment.
RF emissions CISPR11	Class B	The mode Woodpex III is suitable for used in domestic establishment and in establishment directly connected to a low voltage power supply network which supplies buildings used for domestic purposes.
Harmonic emissions IEC 61000-3-2	Class A	
Voltage fluctuations / flicker emissions IEC 61000-3-3	Complies	

<b>Guidance &amp; Declaration — electromagnetic immunity</b>			
The model Woodpex III is intended for use in the electromagnetic environment specified below. The customer or the user of the model Woodpex III should assure that It is used in such an environment.			
<b>Immunity test</b>	<b>IEC 60601 test level</b>	<b>Compliance level</b>	<b>Electromagnetic environment - guidance</b>
Electrostatic discharge (ESD) IEC 61000-4-2	±6 kV contact ±8 kV air	±6 kV contact ±8 kV air	Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30 %.
Electrical fast transient/burst IEC 61000-4-4	±2kV for power supply lines ±1 kV for Input/output lines	±2kV for power supply lines ±1kV for interconnecting cable	Mains power quality should be that of a typical commercial or hospital environment.
Surge IEC 61000-4-5	±1 kV line to line ±2 kV line to earth	±1 kV line to line	Mains power quality should be that of a typical commercial or hospital environment.
Voltage dips, short interruptions and voltage variations on power supply input lines IEC 61000-4-11.	<5 % $U_T$ (>95% dip in $U_T$ ) for 0.5 cycle 40 % $U_T$ (60% dip in $U_T$ ) for 5 cycles 70% $U_T$ (30% dip in $U_T$ ) for 25 cycles <5% $U_T$ (>95 % dip in $U_T$ ) for 5 sec	<5 % $U_T$ (>95% dip in $U_T$ ) for 0.5 cycle 40 % $U_T$ (60% dip in $U_T$ ) for 5 cycles 70% $U_T$ (30% dip in $U_T$ ) for 25 cycles <5% $U_T$ (>95 % dip in $U_T$ ) for 5 sec	Mains power quality should be that of a typical commercial or hospital environment. If the user of the model Woodpex III require continued operation during power mains interruptions, it is recommended that the model Woodpex III be powered from an uninterruptible power supply or a battery.
Power frequency (50/60 Hz ) magnetic field IEC 61000-4-8	3 A/m	3 A/m	Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment.
NOTE $U_T$ is the a.c. mains voltage prior to application of the test level.			

Guidance & Declaration - Electromagnetic immunity			
The model Woodpex III is intended for use in the electromagnetic environment specified below. The customer or the user of the model Woodpex III should assure that it is used in such an environment.			
Immunity test	IEC 60601 test level	Compliance level	Electromagnetic environment - guidance
Conducted RF IEC 61000-4-6 Radiated RF IEC 61000-4-3	3 Vrms 150 kHz to 80 MHz 3 V/m 80 MHz to 2.5 GHz	3V  3 V/m	<p>Portable and mobile RF communications equipment should be used no closer to any part of the model Woodpex III, including cables, than the recommended separation distance calculated from the equation applicable to the frequency of the transmitter.</p> <p><b>Recommended separation distance</b></p> <p>3V  <math>d=1.2 \times P^{1/2}</math> 80 MHz to 800 MHz  <math>d=2.3 \times P^{1/2}</math> 800 MHz to 2.5 GHz</p> <p>where <math>P</math> is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and <math>d</math> is the recommended separation distance in meters (m).</p> <p>Field strengths from fixed RF transmitters, as determined by an electromagnetic site survey,<sup>a</sup> should be less than the compliance level in each frequency range.<sup>b</sup></p> <p>Interference may occur in the vicinity of equipment marked with the following symbol:</p> 
<p>NOTE 1 At 80 MHz and 800 MHz, the higher frequency range applies.</p> <p>NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.</p>			
<p><sup>a</sup> Field strengths from fixed transmitters, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast and TV broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the model Woodpex III is used exceeds the applicable RF compliance level above, the model Woodpex III should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the model Woodpex III.</p> <p><sup>b</sup> Over the frequency range 150 kHz to 80 MHz, field strengths should be less than 3V/m.</p>			

**Recommended separation distances between  
portable and mobile RF communications equipment and the model Woodpex III**

The model Woodpex III is intended for use in electromagnetic environment in which radiated RF disturbances is controlled. The customer or the user of the model Woodpex III can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communications equipment (transmitters) and the model Woodpex III is recommended below, according to the maximum output power of the communications equipment.

Rated maximum output power of transmitter W	Separation distance according to frequency of transmitter m		
	150kHz to 80MHz $d=1.2 \times P^{1/2}$	80MHz to 800MHz $d=1.2 \times P^{1/2}$	800MHz to 2,5GHz $d=2.3 \times P^{1/2}$
0,01	0.12	0.12	0.23
0,1	0.38	0.38	0.73
1	1.2	1.2	2.3
10	3.8	3.8	7.3
100	12	12	23

For transmitters rated at a maximum output power not listed above, the recommended separation distance  $d$  in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where  $P$  is the maximum output power rating of the transmitter in watts (W) accordable to the transmitter manufacturer.

NOTE 1 At 80 MHz and 800 MHz. the separation distance for the higher frequency range applies.

NOTE 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Apex locator in the above specified electromagnetic environment, it will be safe, and it can provide the basic properties such as article 1.6.1-1.6.3;

1. Measurement of pulpitis, pulp necrosis, periapical periodontitis and tooth length.
2. Measurement of the tooth length before restoration of post crown.
3. Measurement of the tooth length of transplantation and retransplantation.

**Cautions:**

1. Cautions: User must regard EMC, please install and put in service Woodpex III according to the EMC information provided in the accompanying documents
2. Cautions: Portable and mobile RF communications equipment can affect medical electrical equipment.
3. Use is not specified for the Apex locator Woodpex III of the adapter, measuring wire, file clip may increase the radiation quantity or reduce the interference ability of the Apex locator system. A list of all cables and maximum lengths of cables is as follows, transducers and other accessories with Guilin Woodpecker Medical Instrument Co., Ltd. claims compliance with the requirements of Emission and Immunity. Please use original accessories.

Serial Number	Accessories name	Cable length	Whether shielding
1	adapter	1	No

2	measuring wire	1.7	No
3	file clip	0.2	No

4. Cautions: The use of accessories, transducers and cables other than those specified, with the exception of transducers and cables sold by Guilin Woodpecker Medical Instrument Co., Ltd. as replacement parts for internal components, may result in increased Emissions or decreased Immunity of Woodpex III.

5. Woodpex III should not be used adjacent to or stacked with other equipment and that if adjacent or stacked use is necessary, Woodpex III should be observed to verify normal operation in the configuration in which it will be used.

6. The accessories adapter, battery, measuring wire, file clip of Apex locator Woodpex III may affect the radiation quantity. The original accessories are in compliance with the requirements of the IEC 60601-1-2. Please use original accessories.

Scan and Login website  
for more information



Guilin Woodpecker Medical Instrument Co., Ltd.  
Information Industrial Park, Guilin National High-Tech  
Zone, Guilin, Guangxi, 541004 P. R. China

Tel:

Europe Sales Dept.: +86-773-5873196, +86-773-2125222

North America, South America &

Oceania Sales Dept.: +86-773-5873198, +86-773-2125123

Asia & Africa Sales Dept.: +86-773-5855350, +86-773-2125896

Fax: +86-773-5822450

E-mail: [woodpecker@glwoodpecker.com](mailto:woodpecker@glwoodpecker.com), [sales@glwoodpecker.com](mailto:sales@glwoodpecker.com)

Website: <http://www.glwoodpecker.com>



MedNet GmbH  
Borkstrasse 10 · 48163 Muenster · Germany